

Test report - PMFT 1000 - Breath resistance / Valve test



Operator	Date	Time
Sager	13.04.2021	16:04:37

Basic information

Operation parameter	
Script	Breath Resistance
Aerosol	Pressurized air

Mask information

Mask ID	13214_20210413_160437
Mask los	2060 6310223
Mask type	FFP2 / x / x

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.06 l/min	28.57 Pa	70.00 Pa	✔ pass
Inhalation 2	94.87 l/min	145.49 Pa	240.00 Pa	✔ pass
Exhalation	160.19 l/min	264.57 Pa	300.00 Pa	✔ pass

Calibration valid until 17.08.2021